



## TECHNICAL EVALUATION REPORT OF IPB-450 (Part-II and Part-I)

### ZOOLOGY LAB EQUIPMENTS (PART-II)

Financial Bids shall be opened for technically qualified bidders on 11-06-2018 @ 12:00 P.M for Part-II

Sr. No	Equipments	Qty,	Rays Technology	Scientific Technical Corporation	HS Enterprises	Technically Qualified	Technically Disqualified	Remarks
01	High Performance Micro Centrifuge	01	Model.Z207M Hermel, Germany	Model.mC12 DAIHAN Biomedical, USA	Model.CF-120 Lab Companion Inc,USA	• Scientific Technical Corporation	• Rays Technologies • HS Enterprises	<b>Rays Technologies</b> Authorization Letter is not attached void tender condition No.4. <b>HS Enterprises</b> - The attached Authorization letter is not signed and stamped by the Lab Companion, USA. - Void tender condition No.5 to provide supply order of similar work from other Govt. departments.
02	Digital Water Bath	01	Model.WNB22 Mettler, Germany	Model.Wb-22 DAIHAN Scientific USA	Hb-22 Lab Companion Inc,USA	• Rays Technology • 2.Scientific Technical Corporation	• HS Enterprises	<b>HS Enterprises</b> - The attached Authorization letter is not signed and stamped by the Lab Companion, USA. - Void tender condition No.5 to provide supply order of similar work from other Govt. departments.
03	Horizontal Gel Electrophoresis System	01	Cat no.1011000 Wealtec, USA	Cat no.ME10-7-10 & Mini-300 Major Science, USA	Model.EP-05 & EPS-300 Lab Companion Inc,USA	• 1.Rays Technology • Scientific Technical Corporation	• HS Enterprises	<b>HS Enterprises</b> - The attached Authorization letter is not signed and stamped by the Lab Companion, USA. - Void tender condition No.5 to provide supply order of similar work from other Govt. departments.
04	Thermal Cycler PCR	01	Model. Multi Gene, Labnet, USA	CYCLER-25 Major Science, USA	Model.T-960 Heal Force, USA	• Scientific Technical Corporation	• Rays Technologies • HS Enterprises	<b>Rays Technologies</b> Temperature Range is below than required. <b>HS Enterprises</b> - Authorization Letter is not attached void tender condition No.4. - Void tender condition No.5 to provide supply order of similar work from other Govt. departments.
05	Variable volume single	01	Gilson Micro pipettes	JENCONS Variable volume	Cat,No.BCPC012S			<b>Rays Technologies</b>



**G S C W U**  
**The Government**  
**Sadiq College Women**  
**University Bahawalpur Pakistan**

Sr. No	Equipments	Qty,	Rays Technology	Scientific Technical Corporation	HS Enterprises	Technically Qualified	Technically Disqualified	Remarks
	<b>channel pipette</b>		Made in France	single channel pipette England	Lab Companion Inc,USA			<b>Measuring Range is not matched for category (a).</b> <b>HS Enterprises</b> - The attached Authorization letter is not signed and stamped by the Lab Companion, USA. - Void tender condition No.5 to provide supply order of similar work from other Govt. departments.
	a) Single Channel Volume, µl: 0.1 – 2.5, Increment, µl: 0.05	01	a) Measuring Range 0.2 to 2µl, Accuracy ±0.014 µl, Cat No.F144561, Model P2N	a) Single Channel Volume, µl: 0.1 – 2.5, Increment, µl: 0.05	a) Single Channel Volume, µl: 0.1 – 2.5, Increment, µl: 0.05	<ul style="list-style-type: none"> <li>Scientific Technical Corporation</li> </ul>	<ul style="list-style-type: none"> <li>Rays Technologies</li> <li>HS Enterprises</li> </ul>	
	b) Single Channel Volume, µl: 0.5 – 10, Increment, µl: 0.1	01	b) Measuring Range 0.5 to 10µl, Accuracy ±0.040 µl, Cat No.F144562, Model P10N	b) Single Channel Volume, µl: 0.5 – 10, Increment, µl: 0.1,	b) Single Channel Volume, µl: 0.5 – 10, Increment, µl: 0.1,	<ul style="list-style-type: none"> <li>Rays Technology</li> <li>Scientific Technical Corporation</li> </ul>	<ul style="list-style-type: none"> <li>HS Enterprises</li> </ul>	
	c) Single Channel Volume, µl: 2 – 20, Increment, µl: 0.5	01	c) Measuring Range 2 to 20µl, Accuracy ±0.060 µl, Cat No.F144563, Model P20N	c) Single Channel Volume, µl: 2 – 20, Increment, µl: 0.5	c) Single Channel Volume, µl: 2 – 20, Increment, µl: 0.5	<ul style="list-style-type: none"> <li>Rays Technology</li> <li>Scientific Technical Corporation</li> </ul>	<ul style="list-style-type: none"> <li>HS Enterprises</li> </ul>	
	d) Single Channel Volume, µl: 5 – 50, Increment, µl: 0.5	01	<b>d) Not quoted</b>	d) Single Channel Volume, µl: 5 – 50, Increment, µl: 0.5	d) Single Channel Volume, µl: 5 – 50, Increment, µl: 0.5	<ul style="list-style-type: none"> <li>Scientific Technical Corporation</li> </ul>	<ul style="list-style-type: none"> <li>HS Enterprises</li> </ul>	
	e) Single Channel Volume, µl: 10 – 100, Increment, µl: 1.0	01	e) Measuring Range 10 to 100µl, Accuracy ±0.015 µl, Cat No.F144564, Model P100N	e) Single Channel Volume, µl: 10 – 100, Increment, µl: 1.0	e) Single Channel Volume, µl: 10 – 100, Increment, µl: 1.0	<ul style="list-style-type: none"> <li>Rays Technology</li> <li>Scientific Technical Corporation</li> </ul>	<ul style="list-style-type: none"> <li>HS Enterprises</li> </ul>	
	f) Single Channel Volume, µl: 20 – 200, Increment, µl: 1.0	01	f) Measuring Range 20 to 200µl, Accuracy ±0.030 µl, Cat No.F144565, Model P200N	f) Single Channel Volume, µl: 20 – 200, Increment, µl: 1.0	f) Single Channel Volume, µl: 20 – 200, Increment, µl: 1.0	<ul style="list-style-type: none"> <li>Rays Technology</li> <li>Scientific Technical Corporation</li> </ul>	<ul style="list-style-type: none"> <li>HS Enterprises</li> </ul>	
	g) Single Channel Volume, µl: 100 – 1000, Increment, µl: 5.0	01	g) Measuring Range 100 to 1000µl, Accuracy ±1.5 µl, Cat No.F144566, Model P1000N	g) Single Channel Volume, µl: 100 – 1000, Increment, µl: 5.0	g) Single Channel Volume, µl: 100 – 1000, Increment, µl: 5.0	<ul style="list-style-type: none"> <li>Rays Technology</li> <li>Scientific Technical Corporation</li> </ul>	<ul style="list-style-type: none"> <li>HS Enterprises</li> </ul>	
	h) Single Channel Volume, ml: 1 – 5, Increment, µl: 50.0	01	h) Measuring Range 1000 to 5000µl, Accuracy ±12µl, Model P5000N	h) Single Channel Volume, ml: 1 – 5, Increment, µl: 50.	h) Single Channel Volume, ml: 1 – 5, Increment, µl: 50.	<ul style="list-style-type: none"> <li>Rays Technology</li> <li>Scientific Technical Corporation</li> </ul>	<ul style="list-style-type: none"> <li>HS Enterprises</li> </ul>	
<b>06</b>	<b>Behavioral Data Recorder</b>	01	<b>Not Quoted</b>	<b>Not Quoted</b>	Behavioral data Recorder Model: (Not mentioned) Company: (Not mentioned)	No Bidder Qualified	<ul style="list-style-type: none"> <li>HS Enterprises</li> </ul>	<b>HS Enterprises</b> - Authorization Letter is not attached void tender condition No.4. - No any Make and Model Mentioned. - Void tender condition No.5 to provide supply order of similar work from other Govt. departments.
<b>07</b>	<b>Ultraviolet Lamp</b>	01	Model.ENF-260C Spectronic, USA	Ultraviolet Lamp Model: Cataloge no.KA.TN.4LC Company: Korea Ace Scientific Co. Ltd.	Ultraviolet Lamp	No Bidder Qualified	<ul style="list-style-type: none"> <li>Rays Technology</li> <li>Scientific Technical Corporation</li> <li>HS Enterprises</li> </ul>	<b>Rays Technologies</b> Authorization Letter is not attached void tender condition No.4. <b>Scientific Technical Corporation</b> Authorization Letter is not



The Government  
Sadiq College Women  
University Bahawalpur Pakistan

Sr. No	Equipments	Qty,	Rays Technology	Scientific Technical Corporation	HS Enterprises	Technically Qualified	Technically Disqualified	Remarks
								attached void tender condition No.4. <b>HS Enterprises</b> - Authorization Letter is not attached void tender condition No.4. - No any Make and Model Mentioned. - Void tender condition No.5 to provide supply order of similar work from other Govt. departments.
08	<b>Trinocular LED Microscope with Built in UPS</b>	01	Model.IM-910 Irmco, Germany	Model.CXL Labomed America, USA	Model.Luxeo Labo America,Inc USA			<b>Rays Technologies</b> - Computer not quoted in accessories. - Authorization letter for Camera is not attached. <b>HS Enterprises</b> - Authorization Letter is not attached void tender condition No.4. - Computer not quoted in accessories. - Void tender condition No.5 to provide supply order of similar work from other Govt. departments. <b>Scientific Technical Corporation.</b> - Authorization letter for Camera and Computer is not attached. - Brand name and Specification of Computer System is not provided.
	<b>Accessories Camera</b>	01	Microscopic Camera Company: Meiji Model: HD1600T 16MP	Camera Model: DSC-W810 Company: SONY	Camera Cosmos: Approx. 20.1 MP			
	<b>Accessories Computer System</b>	01	Not Quoted	Computer System Local Arrangement	Not Quoted		<ul style="list-style-type: none"> <li>• Rays Technology</li> <li>• HS Enterprises</li> <li>• Scientific Technical Corporation</li> </ul>	
09	<b>Variable Speed Laboratory Blender</b>	01	Not Quoted	Model.HGB55E WARING,USA	Oster Pro1200	No bidder Qualified	<ul style="list-style-type: none"> <li>• HS Enterprises</li> <li>• Scientific Technical Corporation</li> </ul>	<b>HS Enterprises</b> - Authorization Letter is not attached void tender condition No.4. - Void tender condition No.5 to provide supply order of similar work from other Govt. departments. <b>Scientific Technical Corporation</b> Authorization Letter is not attached void tender condition No.4.
10	<b>Uv-Spectrophotometer Digital</b>	01	Model.CE 7400 Cecil England.	UV-1900 Biotechnology Medical Services K. Canada	UV-1900 BMS. Canada	<ul style="list-style-type: none"> <li>• Scientific Technical Corporation</li> </ul>	<ul style="list-style-type: none"> <li>• Rays Technologies</li> <li>• HS Enterprises</li> </ul>	<b>Rays Technologies</b> Authorization Letter is not attached void tender condition No.4.



**G**  
**S**  
**C**  
**W**  
**U**  
The Government  
Sadiq College Women  
University Bahawalpur Pakistan

Sr. No	Equipments	Qty,	Rays Technology	Scientific Technical Corporation	HS Enterprises	Technically Qualified	Technically Disqualified	Remarks
								HS Enterprises - Authorization Letter is not attached void tender condition No.4. - Void tender condition No.5 to provide supply order of similar work from other Govt. departments.
11	Camera	01	Not Quoted	Not Quoted	D70 Nikon, Japan	No Bidder Qualified	• HS Enterprises	HS Enterprises - Authorization Letter is not attached void tender condition No.4. - Void tender condition No.5 to provide supply order of similar work from other Govt. departments.

2



The Government  
Sadiq College Women  
University Bahawalpur Pakistan

### Technical Evaluation Report (Part-I)

Tender Notice (IPB-450)

Sr. No.	Name of Items	Method	Qty.	Closing Date and Time	Opening Date and Time	Annex No.	Name of Firms participated	Qualified	Disqualified Reason (s)	Name of Firm for Financial Bid Opening	Financial Bid Opening date and time
1							Nasco Traders	Nasco Traders	---	Nasco Traders	11/06/2018 @ 11:30 A.M.
2	A. Sound System for Cheema B. Hall For wireless audio conferencing system	Single stage two envelope	01	21/05/2018 @ 11:00 A.M.	21/05/2018 @ 11:30 A.M.	Part-I	M/S Taha Enterprises	---	<p><u>M/S Taha Enterprises</u></p> <p>1. According to terms and conditions of Tender Notice, sealed tenders should reach the office of the undersigned on or before 21/05/2018 till 11:00 A.M. and opening of bids shall be at the same day at 11:30 A.M. in the presence of the representative of the competitive firms.</p> <p>2. According to Condition No. 3, "late offers will not be considered"</p> <p>Your tender was received on 21/05/2018 at 12:35 P.M. i.e. after the expiry of stipulated time. Hence, the bid was not included in the tender opening process.</p>	---	---

  
Chairperson Purchase Committee

30-5-2018